

TECHNICAL MARKS FOR 551(100336) NEW NAME (1085)- PROVISION OF INFRASTRUCTURE FACILITIES & EQUIPMENT TO THE UPGRADED GPIS AT TIMERGARA, TAKHTBHAI, SWABI, ABBOTTABAD, KOHAT, NOWSHERA TO THE LEVEL OF COLLEGE OF TECHNOLOGY CIVIL TECHNOLOGY EQUIPMENT DATE OF TENDER OPENING 06-08-2020 DAE CIVIL TECHNOLOGY (MECHANICS OF MATERIALS)

Item No.	Name of Items	Quantity	M/S Al-Waqas Associate, Lahore		M/S Electrical Engineering Services, Lahore.		Rizvi and company, Karachi		Kadeer Brothers Lahore		Glow Pak International, Lahore		ESOLS Engineering, Multan		Ali Trading Peshawar		Lowest	Remarks
			Model	Financial Quoted	Model	Financial Quoted	Model	Financial Quoted	Model	Financial Quoted	Model	Financial Quoted	Model	Financial Quoted	Model	Financial Quoted		
1.	UNIVERSAL TESTING MACHINE (UTM) (Computerized).	1	Not Quoted	Not Quoted	WAW-1000C China	4380000	WAW-1000E China Mentioned specs. Compression space = 570 mm, column distance = 570, power supply 4 KW, 3 ph, 380 VAC Bending Span = 50-500	9930375	WEW-1000B China Space for compression = 520 mm, for tensile = 590 mm, power = 2,1 KW, Piston Stroke = 150 mm	2831400	WAW-1000C China	7200000	SMT-MM-71 PAK	5896800	Not Quoted	0	2831400	Kadeer Brothers Lahore
2.	COMBINED ROCKWELL AND BRINELL HARDNESS TESTER...ASTM-E10-66,	1	HBRV-187.5 China	398000	Mentioned in separate form not in combine form as per tender	Mentioned in separate form not in combine form as per tender	Serie 206EX	1955304	Pressure limit is less than 3000KG Not Qualified	Pressure limit is less than 3000KG Not Qualified	Not Quoted	Not Quoted	SMT-MM-72 CHINA	758160	Not Quoted	0	398000	M/S Al-Waqas Associate, Lahore
3.	CHARPY'S IMPACT TESTING APPARATUS, ASTM- E 23	1	Not Quoted	Not Quoted	JB-300B China	774000	Energy is 50j which is less than required Not Qualified	Energy is 50j which is less than required Not Qualified	JB-300 China	590850	JB-300B China	1400000	SMT-MM-73 PAK	336960	CT-29	245000	245000	Ali Trading Peshawar